

EC Certificate
Directive 93/42/EEC Annex V
Production Quality Assurance
Medical Devices

Registration No.: DD 60148746 0001

Report No.: 17041424 012

Manufacturer: Xiamen Senyang
Co., Ltd.
4-5F, No.99 West 2nd Road, Jiu Tian Hu
Jimei District
Xiamen
361000 Fujian
P.R. China

Products:

- Intermittent pneumatic compression units
- Alternating-Pressure Mattress

Replaces Approval, Registration No.: DD 60147457 0001

Expiry Date: 2024-05-26

The Notified Body hereby declares that the requirements of Annex V of the directive 93/42/EEC have been met for the listed products. The above named manufacturer has established and applies a quality assurance system, which is subject to periodic surveillance, defined by Annex V, section 4 of the aforementioned directive. For placing on the market of class IIb and class III devices covered by this certificate an EC type-examination certificate according to Annex III is required.

Effective Date: 2020-07-24

Date: 2020-07-27

Notified Body



TÜV Rheinland LGA Products GmbH - Tillystraße 2 - 90431 Nürnberg
TÜV Rheinland LGA Products GmbH is a Notified Body according to Directive 93/42/EEC concerning medical devices with the identification number 0197.